

# REMANUFACTURED AMSCO® V-PRO® MAX LOW TEMPERATURE STERILIZATION SYSTEM

# APPLICATION

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System is intended for use in the terminal sterilization of properly prepared (cleaned, rinsed and dried) metal and nonmetal medical devices used in Healthcare Facilities. The STERIS developed low pressure and low temperature Sterilization Cycles are suitable for sterilizing medical devices sensitive to heat and moisture.

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System performs three pre-programmed Sterilization Cycles: the **Lumen Cycle** (approximately 55 minutes to complete), the **Non Lumen Cycle** (approximately 28 minutes to complete) and the **Flexible Cycle** (approximately 35 minutes to complete).

This sterilization system using the **Lumen Cycle** can sterilize\* the following:

- 1. Lumened and non-lumened instruments with diffusionrestricted spaces such as the hinged portion of forceps and scissors.
- 2. Medical devices, including single, dual and triple channeled rigid and semi-rigid endoscopes, with the following configurations:
  - Single channeled devices with a stainless lumen that is  $\ge 0.77 \text{ mm} (\sim 1/32^{"})$  internal diameter (ID) and  $\le 500 \text{ mm} (19-11/16")$  in length
  - <u>Dual channeled devices</u> with stainless lumens that are  $\ge 0.77$  mm (~1/32") ID and  $\le 527$  mm (20-3/4") in length
  - <u>Triple channeled devices</u> with stainless lumens that are either:
  - »  $\geq$  1.2 mm (~3/64") ID and  $\leq$  275 mm (10-52/64") in length
  - » ≥ 1.8 mm (~5/64") ID and ≤ 310 mm (12-13/64") in length OR
  - » ≥  $2.8 \text{ mm} (\sim 7/64") \text{ ID and } \le 317 \text{ mm} (12-31/64") \text{ in length}$
- \* Validation testing was conducted for all lumen sizes using a maximum of 20 lumens per load. Hospital loads should not exceed this validated number of lumens. Validation studies conducted using validation load of two instrument trays and two pouches for a total weight of 19.65 lb (9.0 kg).

This sterilization system using the **Non Lumen Cycle** can sterilize<sup>†</sup> non-lumened instruments including non-lumened general medical instruments, non-lumened rigid, semi-rigid and flexible endoscopes and non-lumened instruments with stainless-steel or titanium diffusion-restricted areas such as the hinged portion of forceps or scissors.

<sup>†</sup> Validation studies conducted using validation load of two instrument trays and two pouches for a total weight of 50 lb (23 kg).



(Typical only - some details may vary.)

This sterilization system using the **Flexible Cycle** can sterilize single or dual lumen surgical flexible endoscopes (such as those used in ENT, Urology and Surgical Care) and bronchoscopes in either of the two configurations:

- 1. Two flexible endoscopes with a light cord (if not integral to endoscope) and mat with no additional load.<sup>‡</sup> The flexible endoscopes may contain either:
  - » A single lumen with an inside diameter of  $\geq$  1 mm (~3/64") and a length of  $\leq$  1050 mm (41")
  - » Or two lumens with:
    - One lumen with an inside diameter of ≥ 1 mm (~3/64") and a length of ≤ 990 mm (39")
    - And the other lumen with an inside diameter of ≥ 1 mm (~3/64") and a length of ≤ 850 mm (33")

<sup>‡</sup> The validation studies were conducted with two flexible endoscopes, each packaged into a tray with silicone mat and light cord (if not integral to endoscope).

#### The Selections Checked Below Apply To This Equipment

#### UNIT POWER SUPPLY

- 200-208/230 Vac, 3 phase, 50/60 Hz, 16 Amp MOUNTING
- □ Single Door, Cabinet
- Single Door, Recessed
- Double Door, Recessed With Panels

#### OPTIONS

- □ IQ/OQ/PQ Protocol (P129385-392)
- □ Seismic Tie-Down Kit (P146669-354)\*

#### **REMOTE MONITORING**

ProConnect<sup>®</sup> Technical Support Services (Remote Monitoring, Priority Technical Support, Customer Care Center Access, Equipment Performance Reports). Available in U.S. and Canada only. \* Conforms to the California Code of Regulations.

Item \_\_\_\_\_

#### Location(s)\_

- 2. One flexible endoscope with a light cord (if not integral to endoscope) and mat and additional non-lumened instruments including instruments with diffusion-restricted areas such as the hinged portion of forceps and scissors." The flexible endoscope may contain either:
  - » A single lumen with an inside diameter of  $\geq$  1 mm (~3/64") and a length of  $\leq$  1050 mm (41")
  - » Or two lumens with:
    - One lumen with an inside diameter of ≥ 1 mm (~3/64") and a length of ≤ 990 mm (39")
    - And the other lumen with an inside diameter of ≥ 1 mm (~3/64") and a length of ≤ 850 mm (33")
- \*\* The validation studies were conducted with a flexible endoscope in a tray with silicone mat and light cord (if not integral to endoscope). Also included in the load were an additional instrument tray and one pouch for a total weight of 24 lb (11 kg).

# NOTE: Single-Use, medical grade polyethylene (PE) and Teflon<sup>®1</sup> tubing<sup>††</sup> with internal diameter $\geq$ 1 mm and length $\leq$ 4 m may be sterilized in the Flexible Cycle. Up to 26 pieces of tubing with no additional load may be sterilized.

<sup>1</sup> Teflon is a registered trademark of E. I. duPont de Nemours and Company. <sup>††</sup> Tubing claims have not been reviewed by the US FDA.

## DESCRIPTION

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System is specifically designed to only process goods using vaporized hydrogen peroxide under vacuum conditions. The process is fully automated, is compatible with a broad range of materials and has rapid Sterilization Cycle times. There are no toxic by-products created by the Sterilization Cycle – only water vapor and oxygen are produced.

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System is NOT intended to process liquids, linens, powders or cellulose materials.

The system utilizes specially designed, disposable, multi-use Cartridges/Cups (available separately) containing VAPROX<sup>®</sup> HC Sterilant and is available in both single and double door configurations.

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System allows Customers to utilize all the advantages of the system but at a significant cost savings. The remanufacturing process includes many new components<sup>‡‡</sup>:

- 10.4" Display Screen
- Electrical Panel
- Exterior Panels and Door(s)
- Printer
- Injection Assembly
- Filters
- <sup>##</sup> Components reused from a used sterilization system undergo an inspection process to ensure reliability. Examples of these components: Chamber, Frame, Vacuum Pump.

# STANDARDS

This Remanufactured Amsco V-PRO maX Low Temperature Sterilization System meets applicable requirements of following standards, **as certified by INTERTEK Testing Services**:

- Underwriters Laboratories (UL) Standard UL 61010-1
  Second Edition
- Canadian Standards Association (CSA) CAN/CSA 22.2 No. 61010-1 Second Edition

#### Governing Directive for the affixing of the CE mark:

Medical Device Directive (MDD) 2007/47/EC

# Standards applied to demonstrate conformity to the directives:

- EN 61010-1 EN 60601-1-2 EN 14937
- IEC 61010-2-040

Each sterilization system is designed, fabricated, assembled and tested in accordance with all applicable sections of UL and CSA.

# SIZE (W X H X L)

**Overall Dimensions:** 

 Cabinet: 33 x 75-1/8 x 38-5/16" (838 x 1908 x 973 mm) Recessed: 33 x 75-1/8 x 37-7/8" (838 x 1908 x 962 mm)

Chamber Size:

• 17 x 15 x 32-1/2" (432 x 381 x 826 mm)

Chamber Volume:

• 4.8 cubic feet (136 L)

## **FEATURES**

The chamber and door assembly are aluminum equipped with a silicone rubber gasket for the door and a welded backhead for the chamber.

Insulation fitted on the chamber wall exterior, door and backhead is 1" (25 mm) thick (nominal). Insulation is held in place with hook-and-loop closures.

Insulation is constructed of oil and water resistant (silicone impregnated) fiberglass.

Automatic door locking mechanism keeps the sterilization system door locked during the entire Sterilization Cycle. After cycle completion, air pressure is used to unlock the door. The sterilizer door cannot be opened if either electrical power or air pressure is lost during sterilizer operation. When sterilization system is in Standby mode, there are no door opening restrictions.

**Chamber heating** is achieved through electric strip heaters attached to the chamber sides, bottom wall, door and backhead. Operating temperature is approximately 122°F (50°C)

**Sterilant cartridge interface** only accepts VAPROX HC Sterilant Cartridges (Cups). The system control automatically tracks the amount of VAPROX HC Sterilant used and the Sterilant expiration date. The control prompts the user on the control display when a new Cup is needed.

The proprietary Cup is equipped with a data matrix code to ensure the correct Cup is used in the sterilization unit and that the Cup contents are not expired; no Cup code (or other information) needs to be entered by the user.

There are two Cup options available:

- Ground Shipped (good for approximately 15 Cycles)
- Air Shipped (good for approximately four Cycles)

Once Cup is punctured by the Sterilization System, it has a 14day "in Unit" shelf life. After use, empty Cups may be disposed in regular waste.

**Catalytic converter** receives outflow from chamber during all cycle phases. Catalytic converter converts hydrogen peroxide into water vapor and oxygen.



#### **Other Components:**

The following are furnished to obtain a complete working unit, ready for (but not including) connection to the facility service lines:

- Resistance Temperature Detectors (RTDs) are installed for sensing and displaying temperature control of vaporizer and chamber. Signals from all system RTDs, converted into electrical impulses, provide accurate control inputs and readouts throughout the entire cycle.
- **Pressure Transducers** are installed for sensing and displaying chamber pressure control. Signals from all system pressure transducers, converted into electrical impulses, provide accurate control inputs and readouts throughout the entire cycle.
- Pneumatic and Solenoid Valves and Switches are used in the sterilization system design to simplify piping and increase serviceability.
- Air Supply and Vacuum Filters are supplied to ensure air entering chamber is High Efficiency Particulate Air (HEPA) filtered (prevent chamber recontamination) and air exhausted from vacuum pump is free of entrapped oil and odor.
- Sterilization System Panels are constructed of plastic and stainless steel.
- Sterilization System Frame and support system is constructed of welded carbon steel with protective paint.
- **High Power Vacuum Pump** is supplied to produce cycle vacuum pulses that remove air and moisture from the chamber. The direct drive rotary vane pump is quiet (<60 dB) with low vibration. A powerful 3 HP (2.2 kW) motor produces a displacement of 53 CFM (90 m<sup>3</sup>/hr) and helps alleviate moisture sensitivity in the sterilization unit. The Sterilizer operating pressure is from atmospheric pressure down to less than 1 Torr.

**Connectivity** to instrument tracking systems is available.

**ProConnect® Technical Support Services** maximizes operational efficiencies with secure, Internet-based, real-time equipment monitoring. Data from your equipment is used by STERIS to provide pro-active Customer alert notifications, technical support, and predictive maintenance. Online parts ordering, equipment performance dashboards, and online service scheduling at steris.com is also available. (ProConnect Technical Support Services is available in U.S. and Canada only.) Refer to Tech Data sheet *SD983, PROCONNECT TECHNICAL SUPPORT SERVICES*, for details.



# **CONTROL DESCRIPTION**

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System is equipped with an Allen-Bradley Compact Logix<sup>™</sup> (Panel View Plus<sup>™</sup> 6 1000 Display<sup>1</sup>) control system and an impact printer.

- <sup>1</sup> CompactLogix<sup>™</sup> and PanelView<sup>™</sup> Plus 6 1000 are trademarks of Allen-Bradley, a Rockwell Automation Company.
- Control Display Panel is located on the front of the sterilization unit in the center while facing the unit. This color touch panel provides user information and allows user inputs. The display is a 640 x 480 pixel resolution, 10.4" screen. Use of this panel and associated screens is normally self-explanatory. The screens are color coded for operator convenience as follows:
  - » Control Screens:
    - » Condition Phase Green
    - » Sterilize Phase Blue
    - » Aeration Phase Violet
  - » Service Screens Light Blue
  - » Option Screens Purple
  - » Alarm Screens Red

# NOTE: This Sterilization System permits no manual control of the Sterilization Cycles.

The Ready, Status, Standby and Cup Empty screens include a cup level indicator (similar to a cell phone battery indicator) in the lower right corner (see Typical Operator Screens). For normal operation, each bar represents approximately four cycles remaining (e.g., four bars means cup contains enough sterilant for 13-16 cycles).

 Printer is located on either the front of the sterilization unit on the right side while facing the unit (single door) or the non-operating end (double door). This alphanumeric impact printer provides an easy to read permanent record of the Sterilization Cycle. Printer provides a 2-1/4" (5.7 mm), 24-character wide cycle tape and paper takeup.

### **CYCLE DESCRIPTION**

The Remanufactured Amsco V-PRO maX Low Temperature Sterilization System is equipped with three pre-programmed Sterilization Cycles: Lumen Cycle (approximately 55 minutes to complete), Non Lumen Cycle (approximately 28 minutes to complete) and Flexible Cycle (approximately 35 minutes to complete). Each Sterilization Cycle proceeds through three phases: CONDITION, STERILIZE and AERATION.

 CONDITION — This cycle phase is a set time vacuum pulse to remove air and moisture from the chamber. When setpoint is reached, load is tested for acceptable moisture content. If content is acceptable, cycle proceeds. If not, Condition pulse repeats.

NOTE: If Condition phase fails the second moisture check, the cycle Aborts.

- **STERILIZE** This cycle phase is a series of four pulses. Each pulse consists of: vacuum pulled to setpoint; VAPROX HC Sterilant vapor drawn into chamber; hold for programmed time; filtered air is introduced to setpoint; hold for programmed time; deep vacuum pulled to setpoint.
- **AERATION** This cycle phase pulls a vacuum to setpoint and continues to evacuate for programmed time to reduce chamber vapor concentration. Once Aeration phase is complete, chamber pressure returns to atmospheric and chamber door is unlocked.

#### **PREVENTIVE MAINTENANCE**

Customers are encouraged to contact STERIS concerning annual maintenance programs. Preventive maintenance, adjustments and replacement of worn parts are provided on a scheduled basis to help ensure optimal equipment performance and help minimize untimely and costly interruptions. STERIS maintains a worldwide staff of wellequipped, factory-trained technicians to provide these services, as well as on-site installation, training, and expert repair services. Contact STERIS for details.

Some of the Amsco V-PRO maX Low Temperature Sterilization System preventive maintenance is usage-based. Vacuum pump preventive maintenance is due every 750 cycles or six months. An oil can (yellow) appears on the screen when it is almost due for maintenance. Can changes to red when maintenance is required (see sample screen).

#### CUSTOMER IS RESPONSIBLE FOR COMPLIANCE WITH APPLICABLE LOCAL AND NATIONAL CODES AND REGULATIONS.

#### NOTES

- 1. For general installation information, refer to STERIS General Notes for Sterilizers (drawing 062941-091). This drawing should always accompany the equipment drawing.
- Refer to equipment drawing showing all utility and space requirements for actual installation specifications. Clearances shown are minimum required for servicing equipment. Floor surface must be hard and level.
- 3. Sterilization system can be installed recessed or with a cabinet package. Each sterilizer is equipped with height adjusting leveling screws. Locking casters are available for installations with limited service clearance. The supplied power cord is six feet; ensure sterilizer is installed within six feet of facility outlet. This sterilizer needs a NEMA L15-20R electrical receptacle.
- 4. STERIS recommends maintaining and operating sterilization system in area where temperature does not exceed 104°F (40°C). Per AAMI ST58:2013, it is also recommended to have a ventilation system exchanging area air at least 10 times per hour.
- 5. STERIS recommends a dedicated, grounded electrical circuit be provided for each unit. Use of an extension cord is not recommended.
- 6. Consult SDS regarding storage and handling of VAPROX HC Sterilant Cartridges (Cups).
- 7. Unit weight 850 lb (386 kg) recessed and 1020 lb (463 kg) free-standing.
- Heat loss at 70°F (21°C) Peak = 1,046 BTU/hr; Avg. = 942 BTU/hr.
- 9. Requirements for Proconnect Services, refer to Tech Data sheet *SD983, PROCONNECT TECHNICAL SUPPORT SERVICES*.
- 10. Electrical Consumption, per cycle = 2.2 kW-hr average; out of cycle = 0.7 kW-hr average.
- 11. STERIS assumes no responsibility for changes to the Remanufactured Amsco V-PRO maX Low Temperature Sterilization System made necessary through failure to observe the supplied necessary specifications (e.g., incorrect facility power supply). Specifications are subject to change without notice.



Sample Preventive Maintenance Screen







# For Further Information, contact:



STERIS Corporation 5960 Heisley Road Mentor, OH 44060-1834 • USA 440-354-2600 • 800-548-4873 www.steris.com